

## DECLARATION OF CONFORMITY UN38.3

We herewith confirm that each battery of this type is proved to meet the requirements of applicable tests in the UN Manual of Tests and Criteria ST/SG/AC.10/11/Rev.5/Amend.1, Part III, Sub-Section 38.3.

In the following, lithium battery test summary according to Sub-Section 38.3.5

- (a), (b) Manufacturer: BMZ GmbH  
Zeche Gustav 1  
63794 Karlstein am Main  
Germany  
Phone: +49 (0) 6188-9956-0  
Fax: +49 (0) 6188-9956-900  
E-Mail: [kontakt@bmz-group.com](mailto:kontakt@bmz-group.com)  
Webpage: [www.bmz-group.com](http://www.bmz-group.com)
- Customer: WEINMANN Emergency Medical  
Technology GmbH + Co.KG  
Fröhösestraße 12  
22525 Hamburg  
Germany  
Phone: +49 (0) 40-88 18 96 0  
Fax: +49 (0) 40-88 18 96 482  
E-Mail: [info@weinmann-emt.de](mailto:info@weinmann-emt.de)  
Webpage: [www.weinmann-emergency.de](http://www.weinmann-emergency.de)
- (c) Test laboratory: Batteryuniversity GmbH  
Am Sportplatz 30  
63791 Karlstein am Main  
Germany  
Phone: +49 (0)6188 – 99410-0  
Fax: +49 (0)6188 – 99410-20  
E-Mail: [mail@bu-lab.eu](mailto:mail@bu-lab.eu)  
Webpage: [www.bu-lab.eu](http://www.bu-lab.eu)
- (d) Report reference no.: BU-2013-000485-UN
- (e) Date of test report: July 31, 2013
- (f) Description of devices under test:
- |                            |  |
|----------------------------|--|
| (i) Type:                  | Lithium-ion batteries                          |
| (ii) Mass:                 | 595 g  |
| (iii) Watt-hour rating:    | 72.90 Wh                                       |
| (iv) Physical description: | 3S3P ICR 18650S3, 10.80 V rechargeable battery |
| (v) Model numbers:         | #9000-3  |
- (g) Performed tests:
- |                             |               |
|-----------------------------|---------------|
| T.1 Altitude simulation:    | Passed        |
| T.2 Thermal test:           | Passed        |
| T.3 Vibration:              | Passed        |
| T.4 Shock:                  | Passed        |
| T.5 External short circuit: | Passed        |
| T.6 Impact/Crush:           | Not performed |
| T.7 Overcharge:             | Passed        |
| T.8 Forced discharge:       | Not performed |
- (h) Reference to assembled battery testing requirements: Not applicable
- (i) Applied standard: UN ST/SG/AC.10/11/Rev.5, Amend.1  
Recommendations of the TRANSPORT OF DANGEROUS GOODS, Manual of Tests and Criteria, Part III, section 38.3, Lithium metal and lithium ion batteries

(j) Signature:

October 28, 2019

Klaus Heck, Managing Director